

ELEKTROTECHNICKÝ ZKUŠEBNÍ ÚSTAV



ELECTROTECHNICAL TESTING INSTITUTE - CZECH REPUBLIC
ELEKTROTECHNISCHE PRÜFANSTALT - TSCHHECHISCHE REPUBLIK
INSTITUT ELECTROTECHNIQUE D'ESSAIS - RÉPUBLIQUE TCHÉQUE
ЭЛЕКТРОТЕХНИЧЕСКИЙ ИСПЫТАТЕЛЬНЫЙ ИНСТИТУТ - ЧЕШСКАЯ РЕСПУБЛИКА

Pod Lisem 129, 171 02 Praha 8 - Troja

EC DESIGN-EXAMINATION CERTIFICATE

issued in accordance with Annex 2 of Government Order No. 54/2015 Coll.
(Annex II of Directive 93/42/EEC)

No.: MED 090046

The Electrotechnical Testing Institute, Notified Body No. 1014, performed the design examination of medical device

**Chirlac braided, Chirlac rapid braided, C-TEC Alfatec braided,
Sterile absorbable synthetic suture in atraumatic needle or without needle**

in accordance with Annex 2 clause 8 of Government Order No. 54/2015 Coll., which specifies technical requirements for medical devices
(Annex II clause 4 of Directive 93/42/EEC) at

manufacturer **CHIRANA T. Injecta, s.r.o.**
Komořanská 2148, 143 00 Praha 4 - Modřany, Czech Republic

and states that the design of medical device meets the provisions of Government Order No. 54/2015 Coll. (Directive 93/42/EEC)

The details of the medical device design examination are presented in the audit report No. 401112-02/01 of 29.4.2014, 401112-02/04,
401112-02/05 of 29.4.2014.


The manufacturer must inform the notified body about any intention of substantial changes to the approved design of medical device which
could affect the conformity with essential requirements in accordance with Annex 1 of Government Order No. 54/2015 Coll. (Annex I of
Directive 93/42/EEC). In that case the conditions under which the certificate has been issued are violated, the notified body may suspend the
validity of the certificate or cancel the certificate.

Edition 2

The first issue of this certificate from 27.05.2009 with validity until 30.04.2014
The validity of this Certificate is limited until: 29.04.2019

30.04.2014 modification 18.09.2014
change 16.03.2016

Prague


Mgr. Miroslav Sedláček
Head of Certification Body



Stamp



401112-02